



MERCADO NACIONAL

PROPOSTA COMERCIAL Nº 0072/2020

DATA: 04/05/2020

BIOTECNOLOGIA

mbbiotecnologia@gmail.com

A Gerencia de Compras

VENDEDOR	ENTREGA	REQUISITANTE	ENVIADO VIA	VALIDADE	PAGAMENTO
MB Biotecnologia	Conforme Proposta		E-MAIL	10 Dias	ANTECIPADO NO PEDIDO

ITEM	DESCRIÇÃO	CODIGO	QUANTIDADE	PREÇO UNITÁRIO	TOTAL
001	Kit IgM/IgG Diagnostic Kit for IgM/IgG Antibody to Coronavirus (SARS-CoV-2), 10 Test Sets/ Kit 20 Test Sets/Kit. Prazo de Entrega: Imediato Fabricante; LIVZON Diagnostic Pagamento: Antecipado	20020	01	170,00	170,00
SUBTOTAL					170,00
TOTAL DEVIDO					170,00

FATURAMENTO MINIMO: 50.000 Kits

FRETE FOB

Dados Bancários:

BANCO DO BRASIL

AGÊNCIA: 1898-8

CONTA CORRENTE: 35856-8

MB Biotecnologia Comercio e Servicos

;

***CANCELAMENTO DE PEDIDO: SOMENTE SERÁ ACEITO ATÉ 48 HORAS APÓS A CONFIRMAÇÃO.**

MB Biotecnologia Comercio e Servicos
(Evandra Regina dos Santos Ramos-ME)
RUA DR. LUIZ MIGLIANO, 2050 – SALA 44 – JD. CABORÉ
SÃO PAULO/SP – CEP. 05711-001
TEL/FAX (11)97319-9656
CNPJ. 14.738.325/0001-80
INSCRIÇÃO ESTADUAL 119.008.709.113
Equipamentos, Reagentes, Plásticos, Assistência Técnica
MERCADO NACIONAL / IMPORTAÇÃO DIRETA



丽珠医药
LIVZON

Diagnostic Kit for IgM / IgG Antibody Coronavirus



For Better Control the Pandemic Outbreak of Novel Co

LIVZON

**Diagnostic Kit for IgM/IgG Antibody to Coronavirus
(Lateral Flow)**



CFDA (NMPA)

FSC



WHO

Whole Blood, Serum and Plasma Sample Types

Result in 1-15min visually


Independent IgM and IgG results



Rapid and Easy to Use, Auxiliary test for the diagnosis

- Ideal complement to coronavirus nucleic acid tests (RT-PCR etc.) to avoid
- Independent IgM and IgG results in one kit, capable for confirmation
(according to the *Diagnosis and Treatment Protocol for Novel Coronavirus Pneumonia 7th Edition* by the National Health Commission)
- Much easier and faster way to test people on site
- Possible to find asymptomatic carriers and to confirm

One-Step Rapid Test for Diagnosing COVID-19



Abundant Clinical Validation Shows Excellent Diagnostic Performance

Tests		Clinical Diagnosis Criteria	
		Diagnosed	Excluded
IgM/IgG	Positive	259	3
	Negative	27	355
Total		286	358

Diagnostic Performance Evaluation and Clinical Validation Data

Diagnostic Kit for IgM/IgG Antibody to Coronavirus (SARS-CoV-2) (Late)

Livzon Diagnostics Inc. March 2020

<i>IgM Antibody to Coronavirus (SARS-CoV-2)</i>	
Test description	<p><i>The product is intended for qualitative detection of Coronavirus IgM antibody in human serum, plasma or whole blood samples. This test is an auxiliary diagnosis method for early infection of Coronavirus (SARS-CoV-2)</i></p> <p><i>Sample type : serum, plasma or whole blood samples</i></p> <p><i>Pack Size: 20T&40T</i></p> <p><i>Storage Condition: 2C°~ 30 C°</i></p> <p><i>Material Provided:</i></p> <p><i>20T:</i></p> <p><i>1.20 Individual sealed pouches, each pouch contains:</i></p> <ul style="list-style-type: none"><i>-1 Tests cassette</i><i>-1 Desiccant pouch</i> <p><i>2.Pipette : 20pcs</i></p> <p><i>Sample diluent : 1*3 mL</i></p> <p><i>40T:</i></p> <p><i>1.40 Individual sealed pouches, each pouch contains:</i></p> <ul style="list-style-type: none"><i>-1 Tests cassette</i><i>-1 Desiccant pouch</i> <p><i>2.Pipette : 40pcs</i></p> <p><i>Sample diluent : 2*3 mL</i></p>
Analytical performance	<p><i>1. Sensitivity and Specificity</i></p> <p><i>Sensitivity: 78.7% (95%CI: 73.6% ~ 83.0%)</i></p> <p><i>Specificity: 99.7% (95%CI: 98.4% ~ 100.0%)</i></p> <p><i>Total agreement: 90.4% (95%CI: 87.8% ~ 92.4%)</i></p> <p><i>2. limit of detection (LOD)</i></p> <p><i>We used the dilution matrix to dilute a specific titer positive control in different series. The results show that when the antibody concentration of 2019-ncov novel coronavirus in the samples was $\geq 1:256$, the detection rate of the reagent was $\geq 95\%$. So the lowest limit of detection is $1:256$.</i></p>

Enterovirus ,Respiratory syncytial virus and Varicella-zoster antibodies positive sample.

4. Interferences

The test result of the kit do not be interfered with the substance at the following concentration:

<i>Substance</i>	<i>Concentration</i>
<i>Bilirubin</i>	<i>50mg/dL</i>
<i>Hemoglobin</i>	<i>50mg/mL</i>
<i>Triglyceride</i>	<i>500mg/dL</i>
<i>Cholesterol</i>	<i>500mg/dL</i>
<i>Rheumatoid factors</i>	<i>920 IU/mL</i>
<i>Antinuclear antibody (ANA) titer</i>	<i>1:1600</i>
<i>Total IgM</i>	<i>80mg/mL</i>
<i>Total IgG</i>	<i>55mg/mL</i>

5. Precision

1). Within run precision: The negative agreement rate and the positive agreement rate should be 100% all the time.

2). Between run precision: The negative agreement rate and the positive agreement rate should be 100% all the time.

6. Reference assay

The reference method used during our clinical study is based on the diagnosis of a diagnosed positive patient that a combination of clinical symptoms and laboratory test was given.

Clinical performance

1. A total of 648 serum/plasma samples from 644 patients with coronavirus pneumonia were tested, including homologous and continuous monitoring samples. After removing 4 samples, a total of 644 samples were included for statistics. The results are as follows:

Compared with the clinical reference standard, the clinical sensitivity was 78.7% (95% CI:73.6% ~ 83.0%). The clinical specificity was 98.4% (CI:98.4% ~ 100.0%), and the total coincidence rate was 92.4%.

2. The combined statistical analysis of the combined detection of the kit and IgG gold standard kit showed that the combined sensitivity was 90.2%, (95% CI:86.2% ~ 93.1%). The combined specificity was 97.6% ~ 99.7%), and the overall coincidence rate was 95.96.6%.

IgG Antibody to Coronavirus (SARS-CoV-2)

Test description

The product is intended for qualitative detection of Coronavirus IgG antibody in human serum, plasma or whole blood samples as an auxiliary diagnosis method for early infection of Coronavirus. Sample type : serum, plasma or whole blood samples.
Pack Size: 20T&40T
Storage Condition: 2C°~ 30 C°
Material Provided:
20T:
1.20 Individual sealed pouches, each pouch contains:
-1 Tests cassette
-1 Desiccant pouch
2.Pipette : 20pcs
Sample diluent : 1*3 mL
40T:
1.40 Individual sealed pouches, each pouch contains:
-1 Tests cassette
-1 Desiccant pouch
2.Pipette : 40pcs
Sample diluent : 2*3 mL

Analytical performance

1. Sensitivity and Specificity
Sensitivity: 83.6% (95% CI: 78.8% ~ 87.4%)
Specificity: 99.4% (95% CI: 98.0% ~ 99.8%)
Total agreement: 92.4% (95% CI:90.1% ~ 94.2%)
2. limit of detection (LOD)
We used the dilution matrix to dilute a specific titer positive control in different series. The results show that when the antibody concentration of 2019-ncov novel coronavirus in the samples was $\geq 1:128$, the detection rate of the reagent was $\geq 95\%$. So the lowest limit of detection of the kit is determined as no less than 1:128.
3. Cross-reactivity
Specimens which tested positive with following various age groups were investigated with the kit. The results showed no cross-reactivity with Endemic human coronaviruses (HKU1, OC43, NL63 and 229E), Influenza B, Nasal virus, EB virus, Mumps virus, Measles virus, Mycoplasma pneumoniae, Chlamydia pneumonia, Adenovirus, Enterovirus, Respiratory syncytial virus and Varicella-zoster virus.

IgG Antibody to Coronavirus (SARS-CoV-2)

Test description

The product is intended for qualitative detection of Coronavirus IgG antibody in human serum, plasma or whole blood samples as an auxiliary diagnosis method for early infection of Coronavirus. Sample type : serum, plasma or whole blood samples.
Pack Size: 20T&40T
Storage Condition: 2C°~ 30 C°
Material Provided:
20T:
1.20 Individual sealed pouches, each pouch contains:
-1 Tests cassette
-1 Desiccant pouch
2.Pipette : 20pcs
Sample diluent : 1*3 mL
40T:
1.40 Individual sealed pouches, each pouch contains:
-1 Tests cassette
-1 Desiccant pouch
2.Pipette : 40pcs
Sample diluent : 2*3 mL

Analytical performance

1. Sensitivity and Specificity
Sensitivity: 83.6% (95% CI: 78.8% ~ 87.4%)
Specificity: 99.4% (95% CI: 98.0% ~ 99.8%)
Total agreement: 92.4% (95% CI:90.1% ~ 94.2%)
2. limit of detection (LOD)
We used the dilution matrix to dilute a specific titer positive control in different series. The results show that when the antibody concentration of 2019-ncov novel coronavirus in the samples was $\geq 1:128$, the detection rate of the reagent was $\geq 95\%$. So the lowest limit of detection of the kit is determined as no less than 1:128.
3. Cross-reactivity
Specimens which tested positive with following various age groups were investigated with the kit. The results showed no cross-reactivity with Endemic human coronaviruses (HKU1, OC43, NL63 and 229E), Influenza B, Nasal virus, EB virus, Mumps virus, Measles virus, Mycoplasma pneumoniae, Chlamydia pneumonia, Adenovirus, Enterovirus, Respiratory syncytial virus and Varicella-zoster virus.

IgM/IgG

Diagnostic Kit for IgM/IgG Antibody to Coronavirus (SARS-CoV-2) (Lateral Flow)

Model: 10 Test Sets/Kit 20 Test Sets/Kit

[Product Name]

Generic Name: Diagnostic Kit for IgM/IgG Antibody to Coronavirus (SARS-CoV-2) (Lateral Flow)

[Intended Use]

This product is used for in vitro qualitative detection of Coronavirus (SARS-CoV-2) IgM/IgG antibody in human serum, plasma and whole blood samples.

[Test Principle]

This product contains one IgM test cassette and one IgG test cassette. The IgM test uses colloidal gold immunochromatography technology to detect the Coronavirus (SARS-CoV-2) IgM antibody. The anti-human IgM antibody labeled by colloidal gold is coated on the conjugate pad. Meanwhile, on the cellulose nitrate membrane, the antigens of Coronavirus (SARS-CoV-2) are coated in the test area, and goat anti-mouse antibody is coated in the control area. While testing, the IgM antibody in the test sample (the specific IgM antibody to Coronavirus (SARS-CoV-2) and non-specific IgM antibody) binds with the colloidal gold labeled anti-human IgM antibody and forms an immune complex. The immune complex moves along the membrane through chromatography. If there is specific IgM antibody to Coronavirus (SARS-CoV-2) in the test sample, it will be captured by the antigens coated in the test area, forming a visible band (test line); The free colloidal gold conjugate or immune complex continues to move forward, and binds specifically with the goat anti-mouse antibody coated in the control area, forming a visible band (control line). If the test sample doesn't contain any Coronavirus (SARS-CoV-2) IgM antibody, only a control line will appear. The IgG test uses colloidal gold immunochromatography technology to detect the Coronavirus (SARS-CoV-2) IgG antibody. The anti-human IgG antibody labeled by colloidal gold is coated on the conjugate pad. Meanwhile, on the cellulose nitrate membrane, the antigens of Coronavirus (SARS-CoV-2) are coated in the test area, and goat anti-mouse antibody is coated in the control area. While testing, the IgG antibody in the test sample (the specific IgG antibody to Coronavirus (SARS-CoV-2) and non-specific IgG antibody) binds with the colloidal gold labeled anti-human IgG antibody and forms an immune complex. The immune complex moves along the membrane through chromatography. If there is specific IgG antibody to Coronavirus (SARS-CoV-2) in the test sample, it will be captured by the antigens coated in the test area forming a visible band (test line).

[Components]

No.	Components
1	IgM test cassette On the cellulose nitrate membrane, the test area is pre-coated with recombinant antigens of the Coronavirus (SARS-CoV-2), while the control area is pre-coated with the goat anti- mouse antibody. The fiberglass is pre-coated with the mouse anti-human IgM antibody labeled by colloidal gold
2	IgG test cassette On the cellulose nitrate membrane, the test area is pre-coated with recombinant antigen of the Coronavirus (SARS-CoV-2), while the control area is pre-coated with the goat anti- mouse antibody. The fiberglass is pre-coated with the mouse anti-human IgG antibody labeled by colloidal gold
3	Pipette
4	Sample diluent buffer solution containing sodium chloride

Note: Components from the different b

[Materials required but not provided]

1. Timer
2. Container for collecting samples

[Storage and shelf life]

1. Store at 2~30°C. The shelf life is 12 months.
2. Keep dry and keep in dark place.
3. The test should be finished within 15 minutes after the foil bag is unsealed. In case of the ambient humidity above 70%, it should be finished within 10 minutes.
4. The manufacture date and expiry date are marked on the kit.

[Sample Collection, handling and storage]

1. The product can be used for testing human serum, plasma and whole blood samples.
2. The plasma or whole blood sample should be mixed with sodium citrate, EDTA-K₂ or heparin.
3. Samples with hemolysis, high viscosity or bacterial contamination are not suitable for testing.
4. Serum or plasma samples can be stored at 2-8°C for short term storage, the samples should be thawed at room temperature after freezing and thawing. The whole blood samples should be tested within 5 days and stored at 2-8°C.

[Test Procedures]


1. Preparation
 - a) Take out the tested samples and the kit from the storage condition and allow them to reach the room temperature.

into the sample well (S) of IgM test cassette and IgG test cassette, and add vertically 2 drops (about 100µL) of sample diluent.

For whole blood sample: add 20µL of whole blood sample into the sample well (S) of IgM test card and IgG test card, and add vertically 2 drops (about 100µL) of sample diluent.

b) Within 1-15 minutes after sample addition, the result can be interpreted as positive while both the control line and test line appear. If only the control line appears and the test line does not appear in 15 minutes, the result can be interpreted as negative. It is invalid to read result after 15 minutes.

[Interpretation of Test Result]



1. Positive result:

a) IgM positive, IgG positive: IgM test cassette appears both control line (C) and test line (T); IgG test cassette appears both control line (C) and test line (T).

b) IgM positive, IgG negative: IgM test cassette appears both control line (C) and test line (T); IgG test cassette appears only control line (C), but no test line (T).

c) IgM negative, IgG positive: IgM test cassette appears only control line (C), but no test line (T); IgG test cassette appears both control line (C) and test line (T).

2. Negative result: IgM test cassette appears only control line (C), but no test line (T); IgG test cassette appears only control line (C), but no test line (T).

3. Invalid result: If IgM and/or IgG test cassette appears no control line (C), no matter whether the test line (T) appears or not, the test result is invalid. A repeat test should be done in case of invalid result appears.

[Limitations of the Test Procedures]

1. The product can only be used for in vitro test of individual's serum, plasma or whole blood samples.

2. A Coronavirus (SARS-CoV-2) infection may not be excluded if the test result is negative.

3. The test result is only for clinical reference and should not be regarded as the only reference for clinical diagnosis and treatment. The clinical management of patients should be considered in combination with their symptoms, physical signs, medical history, other laboratory tests (especially etiology test), treatment response and epidemiological information.

4. In patients with impaired immune function or receiving immunosuppressive therapy, the value of serological antibody test is limited.

5. IgM antibody positive not only occurs in primary infection, but also in secondary infection.

6. The IgM and IgG antibody of Coronavirus (SARS-CoV-2) which this product targets does not directly reflect the presence of Coronavirus

3. The product is intended for qualitative test result.

4. The kit should be used within the shelf life.

5. The test cassettes and pipettes are for single use.

6. Because of the difference in titers of samples, the test results may show different color intensity, all of which are acceptable. The color intensity of the test line cannot be used for determining the antibody titer in the sample.

7. Before testing, the samples stored at low temperature should be allowed to reach the room temperature and be equilibrated for 15 minutes.

8. An inactivation of 56°C incubation for 30 minutes is recommended for the test result.

9. Samples and wastes must be handled according to local regulations and the desiccant in the aluminum foil should be replaced.

[Manufacturer]

Name: Zhuhai Livzon Diagnostics Inc.

Address: No. 266 Tongchang Road, Xianqiao Community, Zhuhai City, Guangdong, P. R. China

Contact: Tel: +86-0756-8919728 Fax: +86-0756-8919728

[EU Representative]

Name: Shanghai International Holding

Address: Eiffelstrasse 80, 20537 Hamburg

Contact: Tel: +49-40-2513175 Fax: +49-40-2513175

Glossary of Symbols



Manufacturer



Authorized Representative



In Vitro diagnostic medical device



Batch Code/Lot Number



Date of Manufacture



Temperature limit



Contains sufficient for <n> Tests



Use By/Expiry Date



CE marking according to IVD Medical Device Regulation



Consult instructions for Use



Sample Diluent



Biological risks



营业执照



码

114A

(副本) (副本号: 5-2)



法定代表人 陶德胜

珠海丽珠试剂股份有限公司(外商投资企业投资, 未上市)

成立日期 1989年01月26日

住所 珠海市香洲区同昌路266号1栋

本营业执照在有效期内有效。经营者应当在每年12月31日前向登记机关报送年度报告。年度报告经登记机关公示无异议的, 方可开展经营活动。经营者应当在每年12月31日前向登记机关报送年度报告。年度报告经登记机关公示无异议的, 方可开展经营活动。经营者应当在每年12月31日前向登记机关报送年度报告。年度报告经登记机关公示无异议的, 方可开展经营活动。

登记机关: 珠海市市场监督管理局
 网址: <http://sggs.zhuhai.gov.cn>



医疗器械生产许可证



名称：珠海丽珠试剂股份有限公司

代表人：陶德胜

负责人：林艳

许可证编号：粤食药监械生产许 20020579 号

生产地址：珠海市香洲区同昌路 266 号

生产范围：见医疗器械生产产品登记表

供货

所：珠海市香洲区同昌路 266 号 1 栋

发证部门：广东省药品监督管理局



医疗器械经营许可证

许可证编号：粤珠食药监械经营许20150120号

企业名称：珠海丽珠试剂股份有限公司

法定代表人：陶德胜

经营方式：批发

企业负责人：林艳

所：珠海市香洲区同昌路266号1栋

经营范围：

III类：6815注射穿刺器械，6824医用激光仪器设备，6833医用核素设备，6840临床检验分析仪器及诊断试剂（诊断试剂不需低温冷藏运输贮存），6840临床检验分析仪器及诊断试剂（诊断试剂需低温冷藏运输贮存），6845体外循环及血液处理设备，6854手术室、急救室、诊疗室设备及器具，6857消毒和灭菌设备及器具，6858医用冷疗、低温、冷藏设备及器具，6870软件**

营场所：珠海市香洲区同昌路266号1栋3层A区



房地址：

珠海市香洲区同昌路266号2栋1层
珠海市香洲区同昌路266号10、珠海市香洲区同

发证部门：

珠海市食品药品监督管理局





中华人民共和国 药品生产许可证

珠海丽珠试剂股份有限公司

珠海市香洲区同昌路266号1栋

91440400617488114A

何德胜

林艳

戴峻英

2020



编号 粤20160266

分类码 T

生产地址和生产范围

珠海市香洲区同昌路266号

体外诊断试剂。

存 档





中华人民共和国

经营许可证



证 号: 粤 AA7560123

经营方式: 批 发

经营范围: 体外诊断试剂**

珠海市香洲区同昌路 266 号 1 栋 3 层

陶德胜

林艳

廖桂芳

珠海市香洲区同昌路 266 号 3 栋 1 层 3101、3102、3103

存

林

品
監

中华人民共和国
药品经营质量管理规范认证证书

证书编号：A-GD-19-0537



珠海丽珠制药股份有限公司
珠海市香洲区同昌路266号1栋3层

存

药品批发

经审查，符合《药品经营质量管理规范》要求，特发此证。





珠海丽珠试剂股份有限公司
ZHUHAI LIVZON DIAGNOSTICS INC.

Declaration of Conformity

according to the In Vitro
 Diagnostic Medical Devices Directive 98/79/EC

Manufacturer: Zhuhai Livzon Diagnostics Inc.
 Address: No.266,Tongchang Road, Xiangzhou District, 519060 Z
 EU Representative: Shanghai International Holding Corp. GmbH (Europe)
 Address: Eiffestrasse 80, 20537 Hamburg, Germany

We declare under our sole responsibility that

the IVD Medical Device	Product Name	:	Diagnostic Kit for IgM/IgG A Coronavirus (SARS-CoV-2) (
	Type/model, batch/serial number, possibly sources and number of items (Where applicable)	:	Model: 10 Test Sets/Kit 20 T
of class	according to directive 98/79/EC	:	Other device ,not in Annex testing ,not for performanc

meets all the provisions of the directive 98/79/EC which apply to it.

Applied harmonised standards, national standards or other normative documents	EN ISO15223-1:2016	EN ISO 13485:2016
	EN 13612:2002	EN 23640:2015
	EN 13641:2002	EN 13975:2003
	EN ISO 14971:2012	EN ISO 18113-1:2011
	EN ISO 18113-2:2011	1272/2008 EC

Conformity assessment procedure

Annex III : EC Declaration of Conformity

符合性声明

按体外诊断医疗器械指令 98/79/EC

制造者名称: 珠海丽珠试剂股份有限公司
地 址: 珠海市香洲区同昌路266号 邮政编码: 519060
欧盟代表: Shanghai International Holding Corp. GmbH (Europe)
地 址: Eiffestrasse 80, 20537 Hamburg, Germany

我方完全以自己的责任声明

体外诊断医
疗器械

名称 : 新型冠状病毒IgM/IgG抗体检测试剂

型号, 批号、系列号,
可能的来源和货号
(如适用) : 规格: 10人份/盒, 20人份/盒

分类

按体外诊断医疗器械
指令 : 非附录 II 中的其他医疗器械, 不
性能评估

满足医疗器械指令98/79/EC中所有适用的条款。

适用的协调化
标准, 国家标
准和其它标准
化文件

EN ISO15223-1:2016
EN 13612:2002
EN 13641:2002
EN ISO 14971:2012
EN ISO 18113-2:2011

EN ISO 13485:2016
EN 23640:2015
EN 13975:2003
EN ISO 18113-1:2011
1272/2008 EC

符合性评审程序 附录 III: EC合格声明

中华人民共和国
PEOPLE'S REPUBLIC OF CHINA
医疗器械产品出口销售证明
CERTIFICATE FOR EXPORTATION OF MEDICAL
PRODUCTS

证书编号：粤食药监械出 20200244 号

Certificate NO.: 粤食药监械出 20200244 号

产品名称：新型冠状病毒（2019-nCoV）IgM/IgG 抗体检测试剂盒（胶体金法
Product (s) : **Diagnostic Kit for IgM/IgG Antibody to Coronavirus (SARS-Flow)**

规格型号：10 人份/盒，20 人份/盒

Model: **10 Test Sets/Kit 20 Test Sets/Kit**

产品注册或备案凭证号：国械注准 20203400240

Registration certificate (s): 国械注准 20203400240

生产企业：珠海丽珠试剂股份有限公司

Manufacturer: **Zhuhai Livzon Diagnostics Inc.**

生产企业住所：珠海市香洲区同昌路 266 号 1 栋

Address of manufacturer: **Building 1, No.266, Tongchang Road, Xiangzhou D**

PEOPLE'S REPUBLIC OF CHINA

生产许可或备案凭证号：粤食药监械生产许 20020579 号

Manufacturing License (s): 粤食药监械生产许 20020579 号

兹证明上述产品已准许在中国生产和销售。

This is to certify that the above products have been registered to manufactured and sold in China.

证明有效日期至：2021 年 03 月 13 日

This certification valid until: 13/03/2021

备注：/

Remark: /





Certificate

Q5 098773 0003 Rev. 00

Holder of Certificate: **Zhuhai Livzon Diagnostics Inc**
No.266,Tongchang Road, Xiangzhou District
519060 Zhuhai
PEOPLE'S REPUBLIC OF CHINA

Facility(ies): Zhuhai Livzon Diagnostics Inc.
No.266,Tongchang Road, Xiangzhou District
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate: **Design and Development,
Production and Distribution of
ELISA Diagnostic Kit,
Colloidal Gold Diagnostic Kit,
Clinical Chemistry Diagnostic Kit,
Microbiology Identification Kit and Me
Instruments for Clinical Laboratory an**

Applied Standard(s): EN ISO 13485:2016
Medical devices - Quality management system
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: SH1829511

TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD
CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFIKAT ◆ CERTIFICAT

cotação kit COVID19

1 mensagem

WR Research Products <wr@wrresearch.com>

4 de maio de 2020 18:30

Para: compras@lge.ibi.unicamp.com.br, compras.csp@unifesp.br, compras.heverton@santamarcelina.org, comprasbertioga@ints.org.br, "Compras - Hospital Dr. Fajardo" <compras.fajardo@saude.am.gov.br>, Setor de Compras <compras@igen.org.br>, anamsilva@hcrp.usp.br, ana.requejo@unisantos.br, ANGELOFRANCA@ints.com.br, silvialima@ints.org.br, GRACIELLEOLIVEIRA@ints.org.br, patricia.uchima@santacasasp.org.br, sildelanhesi@gmail.com, silvanadelanhesi@hsav.com.br, Renato Ferreira De Alencar <renato.alencar@uhgbrasil.com.br>, Thalita Camargo <camargo.thalita@huhsp.org.br>, Thaís Monteiro <thais.monteiro@idor.org>, ctapenna@valinhos.sp.gov.br, fabio.martins@msenior.com.br, hosp_fajardo@saude.am.gov.br, anderson20a@hotmail.com, compra.28agosto@saude.am.gov.br, leticia.costa@santacasasorocaba.com.br, compras.chid@spdm-pais.org.br

Segue cotação kit COVID19, entrega imediata, com registro ANVISA.

Atenciosamente,



BIOTECNOLOGIA

Evandra Ramos

MB Biotecnologia Comercio e Servicos

CNPJ. 14.738.325/0001-80

Rua Dr. Luiz Migliano, 2050 - Sl. 78 – Morumbi

São Paulo/SP - Cep. 05711-001 – Fone (11)973199656 e (11)99970-0017

WhatsApp: 11-99970-0017 - E-mail: Evandra@wrresearch.com

***Por gentileza Confirmar o recebimento desta mensagem!**

2 anexos**PR-072-2020 - COMPRAS.doc**

101K

**COVID.pdf**

9152K